
Midmark M11 - v1.0.0
Total Cycles: 3231

Name: RITTER 1
Sterilizer ID: V2178198
05/15/2023 11:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.8 F 0.0
2:00 183.0 F 0.0
4:00 196.5 F 2.8
6:00 207.5 F 7.8
8:00 222.3 F 13.0
10:00 232.0 F 8.6
12:00 238.8 F 11.0
14:00 243.1 F 12.1
16:00 241.2 F 10.8
18:00 254.7 F 17.6
20:00 266.2 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.5 F 28.8
2:00 271.4 F 28.0
3:00 272.6 F 28.9
4:00 271.8 F 28.3

Min 271.3 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 229.5 F 0.0
10:00 243.6 F 0.0
15:00 238.0 F 0.0
20:00 225.2 F 0.0
25:00 221.7 F 0.0
30:00 221.8 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 21:39
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:46
CYCLE COMPLETE

05/15/2023 12:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____