

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1852

Name: RITTER2  
Sterilizer ID: V2313228  
12/19/2022 09:34 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.1 F 0.0  
2:00 77.4 F 0.0  
4:00 108.4 F 0.0  
6:00 141.2 F 0.0  
8:00 145.0 F 0.1  
10:00 174.9 F 0.7  
12:00 196.5 F 3.6  
14:00 218.1 F 7.1  
16:00 225.5 F 11.2  
18:00 226.3 F 6.6  
20:00 238.2 F 12.1  
22:00 237.0 F 9.8  
24:00 237.3 F 8.9  
26:00 249.3 F 15.1  
28:00 260.9 F 21.7  
30:00 271.6 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.8  
1:00 272.7 F 29.4  
2:00 271.5 F 28.8  
3:00 272.6 F 29.5  
4:00 271.7 F 28.7

Min 271.4 F 28.6  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.1 F 0.0  
5:00 206.5 F 0.0  
10:00 215.5 F 0.0  
15:00 203.2 F 0.0  
20:00 199.8 F 0.0  
25:00 202.2 F 0.0  
30:00 224.1 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 30:38  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:08:41  
CYCLE COMPLETE

12/19/2022 10:42 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_