
Midmark M11 - v1.0.5
Total Cycles: 1973

Name: RITTER2
Sterilizer ID: V2313228
01/25/2023 01:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.7 F 0.0
2:00 150.9 F 0.2
4:00 174.4 F 0.7
6:00 198.6 F 4.5
8:00 216.3 F 9.4
10:00 230.8 F 8.6
12:00 238.1 F 11.2
14:00 235.9 F 9.2
16:00 239.0 F 9.7
18:00 249.9 F 15.3
20:00 261.9 F 22.2
22:00 272.8 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.4 F 29.3
2:00 271.5 F 28.5
3:00 272.4 F 29.5
4:00 272.0 F 28.9

Min 271.2 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.7 F 0.0
5:00 206.9 F 0.0
10:00 223.9 F 0.0
15:00 221.9 F 0.0
20:00 222.4 F 0.0
25:00 227.0 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 22:24
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:46
CYCLE COMPLETE

01/25/2023 02:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____