
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 008444
15:40 24/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:51
PRESSURIZING 2:53
STERILIZING 3:57
136.4 C 223kPa 3:57
136.4 C 224kPa 4:27
136.4 C 225kPa 4:57
136.4 C 225kPa 5:27
136.4 C 223kPa 5:57
136.1 C 220kPa 6:27
136.2 C 223kPa 6:57
136.4 C 223kPa 7:27
136.4 C 223kPa 7:27
VENTING 7:27
AIR DRYING 8:25
CYCLE COMPLETE 20:59

Digital Signature #
90483FC48439E0D1
