
Midmark M11 - v1.0.0
Total Cycles: 3361

Name: RITTER 1
Sterilizer ID: V2178198
06/08/2023 03:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.8 F 0.0
2:00 189.1 F 0.8
4:00 205.7 F 6.1
6:00 221.2 F 11.6
8:00 228.4 F 7.5
10:00 241.1 F 13.3
12:00 241.0 F 11.5
14:00 239.8 F 10.4
16:00 254.1 F 17.5
18:00 266.1 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.7 F 29.0
2:00 271.7 F 28.3
3:00 272.4 F 29.1
4:00 272.4 F 28.8

Min 271.4 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.3 F 0.0
5:00 211.1 F 0.0
10:00 214.6 F 0.0
15:00 221.6 F 0.0
20:00 228.8 F 0.0
25:00 226.5 F 0.0
30:00 220.1 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 19:37
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 00:58:36
CYCLE COMPLETE

06/08/2023 04:15 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____