

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3710

Name: RITTER 1  
Sterilizer ID: V2178198  
08/31/2023 11:14 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.3 F 0.0  
2:00 175.7 F 3.1  
4:00 185.1 F 7.3  
6:00 204.7 F 12.3  
8:00 220.8 F 6.7  
10:00 236.6 F 11.4  
12:00 233.7 F 8.0  
14:00 243.3 F 12.1  
16:00 244.6 F 12.3  
18:00 256.5 F 18.6  
20:00 267.4 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.2 F 28.4  
1:00 272.6 F 28.8  
2:00 272.5 F 29.0  
3:00 271.9 F 28.3  
4:00 272.3 F 28.8  
  
Min 271.7 F 28.3  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.0 F 0.0  
5:00 226.9 F 0.0  
10:00 217.9 F 0.0  
15:00 206.5 F 0.0  
20:00 198.6 F 0.0  
25:00 203.3 F 0.0  
30:00 224.4 F 0.0  
Drying Complete

FILLING: 1:51  
HEATING: 21:58  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:01:14  
CYCLE COMPLETE

08/31/2023 12:15 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_