
Midmark M11 - v1.0.5
Total Cycles: 4067

Name: RITTER2
Sterilizer ID: V2313228
10/30/2024 11:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 112.7 F 0.0
2:00 113.0 F 0.0
4:00 130.3 F 0.1
6:00 161.5 F 1.9
8:00 183.3 F 5.6
10:00 199.6 F 9.8
12:00 215.6 F 8.9
14:00 225.0 F 8.3
16:00 237.3 F 13.1
18:00 236.7 F 10.2
20:00 236.5 F 8.8
22:00 248.4 F 14.9
24:00 259.9 F 21.3
26:00 270.5 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.5 F 29.3
2:00 272.1 F 29.6
3:00 271.9 F 29.0
4:00 272.5 F 29.7

Min 271.4 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.2 F 0.0
5:00 199.9 F 0.0
10:00 188.6 F 0.0
15:00 185.3 F 0.0
20:00 182.0 F 0.0
25:00 180.5 F 0.0
30:00 178.9 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 26:49
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:05:14
CYCLE COMPLETE

10/30/2024 12:14 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____