

-----  
Mississauga, ON  
Canada  
-----

STATIM 5000 S501R751  
SN 511020J00004  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000784  
11:09 10/03/2022  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 5:39  
PRESSURIZING 7:45  
STERILIZING 9:22  
136.0 C 220kPa 9:22  
136.3 C 222kPa 9:52  
136.4 C 220kPa 10:22  
136.0 C 220kPa 10:52  
136.3 C 223kPa 11:22  
136.4 C 223kPa 11:52  
136.4 C 222kPa 12:22  
136.3 C 223kPa 12:52  
136.3 C 223kPa 12:52  
VENTING 12:52  
AIR DRYING 13:59  
CYCLE COMPLETE 28:43

-----  
Digital Signature #  
450A30516A27F690  
-----