
Midmark M11 - v1.0.0
Total Cycles: 4434

Name: RITTER 1
Sterilizer ID: V2178198
04/08/2024 11:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.7 F 0.0
2:00 122.3 F 0.9
4:00 125.7 F 2.8
6:00 149.7 F 5.5
8:00 170.5 F 8.7
10:00 200.2 F 12.6
12:00 209.3 F 4.3
14:00 223.5 F 7.9
16:00 231.0 F 11.9
18:00 231.5 F 7.7
20:00 242.4 F 12.7
22:00 240.1 F 10.6
24:00 251.5 F 16.2
26:00 261.7 F 22.0
28:00 271.3 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.5 F 29.1
2:00 271.7 F 28.6
3:00 271.1 F 28.6
4:00 272.7 F 29.5

Min 271.0 F 28.3
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.2 F 0.0
5:00 207.3 F 0.0
10:00 206.9 F 0.0
15:00 226.4 F 0.0
20:00 216.6 F 0.0
25:00 214.9 F 0.0
30:00 214.3 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 28:44
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:07:25
CYCLE COMPLETE

04/08/2024 12:37 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____