
Midmark M11 - v1.0.5
Total Cycles: 3034

Name: RITTER2
Sterilizer ID: V2313228
10/20/2023 08:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.6 F 0.0
2:00 69.9 F 0.0
4:00 76.4 F 0.2
6:00 95.9 F 1.5
8:00 130.2 F 4.1
10:00 164.4 F 7.6
12:00 187.1 F 11.3
14:00 215.4 F 3.9
16:00 226.9 F 7.5
18:00 234.9 F 11.4
20:00 228.8 F 7.0
22:00 240.8 F 12.3
24:00 238.4 F 10.2
26:00 250.6 F 16.2
28:00 261.5 F 22.3
30:00 271.3 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 272.4 F 30.0
2:00 271.3 F 29.6
3:00 271.9 F 29.5
4:00 272.5 F 29.9

Min 270.9 F 28.9
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.8 F 0.0
5:00 241.0 F 0.0
10:00 222.0 F 0.0
15:00 233.0 F 0.0
20:00 231.4 F 0.0
25:00 222.1 F 0.0
30:00 225.5 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 30:29
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:08:35
CYCLE COMPLETE

10/20/2023 09:17 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____