
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002068
9:48 20/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:32
PRESSURIZING 1:37
STERILIZING 2:42
136.4 C 226kPa 2:42
136.3 C 226kPa 3:12
136.2 C 222kPa 3:42
136.2 C 224kPa 4:12
136.2 C 225kPa 4:42
136.2 C 225kPa 5:12
136.4 C 227kPa 5:42
136.3 C 225kPa 6:12
136.3 C 225kPa 6:12
VENTING 6:12
AIR DRYING 7:03
CYCLE COMPLETE 13:45

Digital Signature #
794F590ACBA70B49
