
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000860
9:08 13/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:13
PRESSURIZING 3:22
STERILIZING 4:34
136.0 C 226kPa 4:34
136.3 C 232kPa 5:04
136.3 C 232kPa 5:34
136.4 C 232kPa 6:04
136.4 C 233kPa 6:34
136.4 C 232kPa 7:04
136.4 C 232kPa 7:34
136.3 C 232kPa 8:04
136.3 C 232kPa 8:04
VENTING 8:04
AIR DRYING 8:57
CYCLE COMPLETE 23:57

Digital Signature #
43228BB3BD8FBB7E
