

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3352

Name: RITTER2  
Sterilizer ID: V2313228  
02/02/2024 03:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 146.4 F 0.0  
2:00 150.8 F 1.8  
4:00 173.9 F 5.3  
6:00 196.5 F 9.6  
8:00 213.5 F 8.4  
10:00 225.0 F 8.2  
12:00 235.7 F 13.4  
14:00 238.6 F 11.0  
16:00 237.6 F 9.7  
18:00 251.7 F 16.3  
20:00 262.5 F 23.0  
22:00 272.9 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.1  
1:00 271.4 F 29.1  
2:00 271.4 F 29.1  
3:00 271.2 F 29.0  
4:00 271.5 F 29.1

Min 271.2 F 29.0  
Max 273.2 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.7 F 0.0  
5:00 205.8 F 0.0  
10:00 201.9 F 0.0  
15:00 200.9 F 0.0  
20:00 198.1 F 0.0  
25:00 196.0 F 0.0  
30:00 194.5 F 0.0  
Drying Complete

FILLING: 1:12  
HEATING: 23:27  
STERILIZING: 4:00  
VENTING: 3:36  
DRYING: 30:00  
TOTAL CYCLE: 01:02:14  
CYCLE COMPLETE

02/02/2024 04:57 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_