
Midmark M11 - v1.0.5
Total Cycles: 1884

Name: RITTER2
Sterilizer ID: V2313228
12/28/2022 08:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.1 F 0.0
2:00 78.3 F 0.0
4:00 81.1 F 0.0
6:00 100.0 F 0.0
8:00 141.5 F 0.1
10:00 174.2 F 2.6
12:00 193.3 F 5.9
14:00 206.6 F 9.7
16:00 219.2 F 11.2
18:00 226.3 F 8.1
20:00 239.7 F 12.9
22:00 239.4 F 10.6
24:00 236.1 F 8.9
26:00 250.0 F 15.5
28:00 261.1 F 21.9
30:00 271.5 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.4
1:00 272.1 F 29.3
2:00 272.4 F 29.5
3:00 272.6 F 29.7
4:00 272.5 F 30.0

Min 271.6 F 29.3
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.2 F 0.0
5:00 208.4 F 0.0
10:00 224.3 F 0.0
15:00 226.3 F 0.0
20:00 229.3 F 0.0
25:00 226.3 F 0.0
30:00 223.9 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:28
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:09:23
CYCLE COMPLETE

12/28/2022 09:48 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____