
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001478
14:23 25/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:37
STERILIZING 3:45
136.0 C 223kPa 3:45
136.4 C 226kPa 4:15
136.3 C 225kPa 4:45
136.3 C 225kPa 5:15
136.3 C 225kPa 5:45
136.3 C 225kPa 6:15
136.4 C 226kPa 6:45
136.4 C 226kPa 7:15
136.4 C 226kPa 7:15
VENTING 7:15
AIR DRYING 8:09
CYCLE COMPLETE 9:11

Digital Signature #
70710A827F0C4382
