

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3054

Name: RITTER2  
Sterilizer ID: V2313228  
10/25/2023 01:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 126.1 F 0.0  
2:00 129.7 F 1.3  
4:00 155.4 F 4.7  
6:00 181.4 F 9.2  
8:00 207.0 F 7.3  
10:00 226.6 F 8.1  
12:00 241.0 F 13.4  
14:00 238.9 F 10.8  
16:00 236.5 F 9.2  
18:00 251.3 F 16.2  
20:00 263.6 F 23.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.8  
1:00 272.6 F 29.4  
2:00 271.1 F 28.7  
3:00 272.6 F 29.5  
4:00 271.8 F 28.9  
  
Min 271.0 F 28.6  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.6 F 0.0  
5:00 248.3 F 0.0  
10:00 219.2 F 0.0  
15:00 215.0 F 0.0  
20:00 208.0 F 0.0  
25:00 203.1 F 0.0  
30:00 204.7 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 22:00  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:00:17  
CYCLE COMPLETE

10/25/2023 02:01 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_