
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002472
9:05 29/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:08
PRESSURIZING 3:17
STERILIZING 4:32
136.2 C 221kPa 4:32
136.1 C 222kPa 5:02
136.4 C 222kPa 5:32
136.2 C 223kPa 6:02
136.4 C 225kPa 6:32
136.4 C 221kPa 7:02
136.4 C 220kPa 7:32
136.4 C 222kPa 8:02
136.4 C 222kPa 8:02
VENTING 8:02
AIR DRYING 8:54
CYCLE COMPLETE 38:54

Digital Signature #
C3511011FE54FE15
