
Midmark M11 - v1.0.0
Total Cycles: 2109

Name: RITTER 1
Sterilizer ID: V2178198
07/08/2022 04:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.6 F 0.0
2:00 179.1 F 3.7
4:00 201.1 F 8.6
6:00 217.4 F 7.8
8:00 233.3 F 9.5
10:00 234.1 F 8.7
12:00 244.0 F 12.8
14:00 242.4 F 11.4
16:00 255.9 F 18.3
18:00 267.5 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.5 F 28.7
2:00 271.8 F 28.4
3:00 272.5 F 28.6
4:00 271.6 F 28.3

Min 271.5 F 28.1
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.7 F 0.0
5:00 226.5 F 0.0
10:00 214.3 F 0.0
15:00 210.2 F 0.0
20:00 215.0 F 0.0
25:00 218.6 F 0.0
30:00 227.1 F 0.0

Drying Complete

FILLING: 2:23
HEATING: 19:19
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:58:46
CYCLE COMPLETE

07/08/2022 05:11 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____