
Midmark M11 - v1.0.5
Total Cycles: 4416

Name: RITTER2
Sterilizer ID: V2313228
03/10/2025 11:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 118.1 F 0.0
2:00 132.7 F 0.7
4:00 124.1 F 2.5
6:00 138.2 F 5.4
8:00 164.6 F 8.9
10:00 184.0 F 12.8
12:00 212.8 F 4.9
14:00 215.2 F 8.6
16:00 226.6 F 12.6
18:00 233.4 F 8.9
20:00 238.9 F 12.2
22:00 242.5 F 12.0
24:00 253.2 F 17.6
26:00 262.8 F 23.5
28:00 272.4 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.7
1:00 272.7 F 30.2
2:00 271.1 F 29.6
3:00 272.1 F 30.3
4:00 271.4 F 29.7

Min 271.0 F 29.6
Max 272.9 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.4 F 0.0
5:00 228.6 F 0.0
10:00 221.3 F 0.0
15:00 229.6 F 0.0
20:00 199.5 F 0.0
25:00 212.9 F 0.0
30:00 202.8 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 29:31
STERILIZING: 4:00
VENTING: 3:54
DRYING: 30:00
TOTAL CYCLE: 01:08:20
CYCLE COMPLETE

03/10/2025 12:29 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____