
Midmark M11 - v1.0.5
Total Cycles: 4288

Name: RITTER2
Sterilizer ID: V2313228
01/23/2025 02:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.1 F 0.0
2:00 145.1 F 2.1
4:00 155.5 F 5.5
6:00 182.6 F 9.8
8:00 213.5 F 6.2
10:00 219.8 F 7.5
12:00 233.1 F 11.8
14:00 231.8 F 8.6
16:00 241.3 F 11.6
18:00 244.0 F 12.9
20:00 255.8 F 19.1
22:00 266.6 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.3
1:00 271.9 F 30.0
2:00 271.8 F 29.3
3:00 271.4 F 29.9
4:00 272.3 F 29.6

Min 271.0 F 29.1
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.4 F 0.0
5:00 211.5 F 0.0
10:00 194.6 F 0.0
15:00 184.5 F 0.0
20:00 182.3 F 0.0
25:00 180.8 F 0.0
30:00 180.0 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 23:32
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:02:26
CYCLE COMPLETE

01/23/2025 03:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____