

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2605

Name: RITTER 1  
Sterilizer ID: V2178198  
11/23/2022 02:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 204.0 F 0.0  
2:00 203.3 F 4.5  
4:00 213.7 F 9.9  
6:00 221.4 F 5.3  
8:00 235.9 F 11.1  
10:00 235.8 F 8.5  
12:00 240.2 F 9.7  
14:00 248.3 F 14.5  
16:00 261.1 F 21.6  
18:00 272.5 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 28.4  
1:00 272.1 F 28.7  
2:00 272.3 F 29.0  
3:00 271.6 F 28.4  
4:00 272.1 F 28.7  
  
Min 271.5 F 28.3  
Max 272.5 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.9 F 0.0  
5:00 222.0 F 0.0  
10:00 236.2 F 0.0  
15:00 231.6 F 0.0  
20:00 230.8 F 0.0  
25:00 230.1 F 0.0  
30:00 224.6 F 0.0  
Drying Complete

FILLING: 2:05  
HEATING: 18:09  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:57:27  
CYCLE COMPLETE

11/23/2022 03:12 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_