

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1462

Name: RITTER2  
Sterilizer ID: V2313228  
09/07/2022 11:20 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 171.5 F 0.0  
2:00 175.1 F 3.1  
4:00 183.6 F 7.2  
6:00 205.4 F 12.4  
8:00 219.5 F 6.8  
10:00 230.5 F 12.1  
12:00 234.8 F 9.1  
14:00 238.6 F 9.8  
16:00 246.6 F 14.2  
18:00 259.0 F 20.9  
20:00 270.0 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.2  
1:00 272.7 F 30.0  
2:00 272.3 F 29.3  
3:00 271.5 F 29.4  
4:00 272.7 F 30.0  
  
Min 271.4 F 29.0  
Max 273.0 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.1 F 0.0  
5:00 214.7 F 0.0  
10:00 222.9 F 0.0  
15:00 225.9 F 0.0  
20:00 224.2 F 0.0  
25:00 221.6 F 0.0  
30:00 222.5 F 0.0  
Drying Complete

FILLING: 1:16  
HEATING: 20:58  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:59:22  
CYCLE COMPLETE

09/07/2022 12:19 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_