
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000679
15:41 21/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:28
PRESSURIZING 3:30
STERILIZING 4:30
136.4 C 224kPa 4:30
136.4 C 225kPa 5:00
136.4 C 225kPa 5:30
136.2 C 223kPa 6:00
136.4 C 224kPa 6:30
136.4 C 225kPa 7:00
136.4 C 225kPa 7:30
136.3 C 222kPa 8:00
136.3 C 222kPa 8:00
VENTING 8:00
AIR DRYING 8:56
CYCLE COMPLETE 9:14

Digital Signature #
825A35C788E795A8
