
Midmark M11 - v1.0.0
Total Cycles: 5227

Name: RITTER 1
Sterilizer ID: V2178198
01/08/2025 01:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 93.5 F 0.0

C103: HEATUP MODE
STOP PRESSED

** ITEMS NOT STERILE **

01/08/2025 01:24 PM