

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000276  
13:13 01/10/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:51  
PRESSURIZING 3:52  
STERILIZING 5:16  
136.4 C 233kPa 5:16  
136.2 C 231kPa 5:46  
136.2 C 233kPa 6:16  
136.2 C 232kPa 6:46  
136.3 C 231kPa 7:16  
136.2 C 233kPa 7:46  
136.3 C 233kPa 8:16  
136.4 C 233kPa 8:46  
136.4 C 233kPa 8:46  
VENTING 8:46  
AIR DRYING 9:54  
CYCLE COMPLETE 19:26  
-----

Digital Signature #  
304BB91276A02D99  
-----