
Midmark M11 - v1.0.5
Total Cycles: 1291

Name: RITTER2
Sterilizer ID: V2313228
07/13/2022 01:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.9 F 0.0
2:00 181.7 F 0.3
4:00 197.8 F 3.7
6:00 219.2 F 8.7
8:00 230.3 F 9.7
10:00 236.0 F 10.9
12:00 235.8 F 9.0
14:00 238.5 F 9.8
16:00 248.5 F 14.9
18:00 260.7 F 21.8
20:00 271.6 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.2 F 29.4
2:00 271.5 F 29.3
3:00 272.3 F 29.3
4:00 271.4 F 29.2

Min 271.3 F 28.8
Max 272.4 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.3 F 0.0
5:00 223.9 F 0.0
10:00 235.1 F 0.0
15:00 231.3 F 0.0
20:00 226.3 F 0.0
25:00 224.5 F 0.0
30:00 223.0 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 20:36
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:57
CYCLE COMPLETE

07/13/2022 02:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____