
Midmark M11 - v1.0.5
Total Cycles: 3168

Name: RITTER2
Sterilizer ID: V2313228
11/29/2023 02:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.0 F 0.0
2:00 141.3 F 1.4
4:00 164.0 F 4.2
6:00 186.7 F 7.8
8:00 202.6 F 11.8
10:00 215.9 F 5.3
12:00 223.2 F 9.7
14:00 231.8 F 9.9
16:00 238.4 F 11.6
18:00 237.5 F 10.1
20:00 250.3 F 16.2
22:00 261.2 F 22.4
24:00 271.2 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.4 F 30.1
2:00 271.9 F 29.4
3:00 271.3 F 29.4
4:00 272.4 F 30.1

Min 271.1 F 29.1
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.1 F 0.0
5:00 209.2 F 0.0
10:00 203.8 F 0.0
15:00 196.8 F 0.0
20:00 194.8 F 0.0
25:00 195.1 F 0.0
30:00 195.4 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 24:46
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:03:32
CYCLE COMPLETE

11/29/2023 03:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____