
Midmark M11 - v1.0.0
Total Cycles: 4218

Name: RITTER 1
Sterilizer ID: V2178198
01/19/2024 03:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.3 F 0.0
2:00 170.9 F 3.4
4:00 193.1 F 7.6
6:00 211.6 F 13.3
8:00 227.8 F 7.4
10:00 235.8 F 12.7
12:00 238.3 F 9.9
14:00 237.0 F 8.7
16:00 249.6 F 14.8
18:00 261.5 F 21.5
20:00 272.3 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.1
1:00 272.9 F 29.1
2:00 271.5 F 28.4
3:00 272.7 F 28.9
4:00 271.8 F 28.7

Min 270.9 F 27.9
Max 273.1 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.3 F 0.0
5:00 236.1 F 0.0
10:00 219.5 F 0.0
15:00 196.7 F 0.0
20:00 184.4 F 0.0
25:00 190.7 F 0.0
30:00 224.7 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 20:30
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 00:59:44
CYCLE COMPLETE

01/19/2024 04:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____