
Midmark M11 - v1.0.0
Total Cycles: 1460

Name: RITTER 1
Sterilizer ID: V2178198
12/15/2021 01:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.3 F 0.0
2:00 183.9 F 3.2
4:00 195.4 F 7.4
6:00 211.7 F 12.5
8:00 222.3 F 7.2
10:00 236.1 F 12.9
12:00 237.4 F 10.3
14:00 236.2 F 8.5
16:00 250.3 F 15.7
18:00 262.4 F 22.7
20:00 272.6 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.7
1:00 272.7 F 29.4
2:00 272.5 F 29.0
3:00 271.9 F 28.8
4:00 273.0 F 29.5

Min 271.7 F 28.6
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 225.2 F 0.0
10:00 229.6 F 0.0
15:00 229.4 F 0.0
20:00 226.2 F 0.0
25:00 226.1 F 0.0
30:00 227.8 F 0.0
Drying Complete

FILLING: 1:52
HEATING: 20:07
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:08
CYCLE COMPLETE

12/15/2021 02:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____