

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1747

Name: RITTER 1  
Sterilizer ID: V2178198  
03/25/2022 11:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.3 F 0.0  
2:00 177.6 F 3.1  
4:00 192.3 F 6.8  
6:00 215.9 F 12.8  
8:00 223.0 F 7.2  
10:00 235.6 F 12.4  
12:00 235.9 F 9.5  
14:00 238.0 F 9.0  
16:00 248.0 F 14.4  
18:00 260.3 F 21.2  
20:00 271.3 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.6  
1:00 272.3 F 29.1  
2:00 271.4 F 28.6  
3:00 272.4 F 29.2  
4:00 271.4 F 28.4

Min 271.3 F 28.4  
Max 272.5 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.2 F 0.0  
5:00 238.9 F 0.0  
10:00 225.3 F 0.0  
15:00 223.1 F 0.0  
20:00 225.2 F 0.0  
25:00 229.9 F 0.0  
30:00 230.7 F 0.0  
Drying Complete

FILLING: 1:28  
HEATING: 20:39  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:59:18  
CYCLE COMPLETE

03/25/2022 12:47 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02