
Midmark M11 - v1.0.0
Total Cycles: 1486

Name: RITTER 1
Sterilizer ID: V2178198
12/29/2021 10:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.0 F 0.0
2:00 188.9 F 3.4
4:00 201.2 F 8.0
6:00 227.5 F 10.1
8:00 230.7 F 9.3
10:00 235.9 F 9.4
12:00 243.5 F 12.6
14:00 241.9 F 11.2
16:00 255.1 F 17.9
18:00 266.5 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 272.5 F 28.9
2:00 271.3 F 28.2
3:00 272.2 F 29.1
4:00 272.2 F 28.6

Min 271.1 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 210.5 F 0.0
10:00 231.2 F 0.0
15:00 229.9 F 0.0
20:00 227.8 F 0.0
25:00 230.1 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 19:35
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:34
CYCLE COMPLETE

12/29/2021 11:52 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____