
Midmark M11 - v1.0.0
Total Cycles: 1155

Name: RITTER 1
Sterilizer ID: V2178198
09/09/2021 10:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.5 F 0.0
2:00 182.8 F 3.1
4:00 204.7 F 7.9
6:00 223.8 F 12.4
8:00 231.5 F 9.0
10:00 237.9 F 10.0
12:00 243.4 F 12.3
14:00 241.2 F 10.9
16:00 255.0 F 17.8
18:00 266.5 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.1 F 28.4
2:00 272.4 F 28.9
3:00 271.4 F 28.1
4:00 272.5 F 28.8

Min 271.3 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.6 F 0.0
5:00 226.6 F 0.0
10:00 215.8 F 0.0
15:00 213.3 F 0.0
20:00 191.8 F 0.0
25:00 177.8 F 0.0
30:00 160.4 F 0.0
Drying Complete

FILLING: 2:33
HEATING: 19:27
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:59:06
CYCLE COMPLETE

09/09/2021 11:43 AM
ACCEPTED: X
REJECTED:
Operator: U02