
Midmark M11 - v1.0.5
Total Cycles: 1292

Name: RITTER2
Sterilizer ID: V2313228
07/13/2022 02:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.4 F 0.0
2:00 195.6 F 0.2
4:00 214.5 F 3.3
6:00 220.9 F 8.2
8:00 227.8 F 11.5
10:00 236.9 F 10.1
12:00 234.6 F 8.8
14:00 241.8 F 11.6
16:00 246.0 F 13.6
18:00 258.1 F 20.4
20:00 269.2 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.4 F 29.6
2:00 271.6 F 29.1
3:00 272.5 F 29.7
4:00 271.7 F 29.0

Min 271.5 F 28.9
Max 272.5 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.1 F 0.0
5:00 224.3 F 0.0
10:00 247.7 F 0.0
15:00 241.9 F 0.0
20:00 228.7 F 0.0
25:00 227.1 F 0.0
30:00 224.5 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 21:04
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:26
CYCLE COMPLETE

07/13/2022 03:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____