
Midmark M11 - v1.0.0
Total Cycles: 5331

Name: RITTER 1
Sterilizer ID: V2178198
02/10/2025 05:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.5 F 0.0
2:00 153.2 F 3.2
4:00 172.6 F 6.9
6:00 196.1 F 11.5
8:00 221.1 F 4.9
10:00 232.6 F 10.0
12:00 232.5 F 7.7
14:00 242.8 F 11.9
16:00 239.7 F 10.0
18:00 253.2 F 16.8
20:00 265.1 F 23.7
22:00 271.6 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.5 F 28.9
2:00 272.1 F 29.0
3:00 272.2 F 28.7
4:00 272.3 F 29.2

Min 271.5 F 28.4
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.1 F 0.0
5:00 224.5 F 0.0
10:00 214.8 F 0.0
15:00 214.0 F 0.0
20:00 258.1 F 0.0
25:00 242.1 F 0.0
30:00 222.4 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 22:34
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:01:12
CYCLE COMPLETE

02/10/2025 06:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____