
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 005057
13:25 02/02/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:55
PRESSURIZING 1:57
STERILIZING 3:02
136.4 C 220kPa 3:02
136.4 C 225kPa 3:32
136.4 C 223kPa 4:02
136.2 C 223kPa 4:32
136.4 C 224kPa 5:02
136.4 C 222kPa 5:32
136.4 C 223kPa 6:02
136.3 C 224kPa 6:32
136.3 C 224kPa 6:32
VENTING 6:32
AIR DRYING 7:25
CYCLE COMPLETE 9:20

Digital Signature #
E60C726C11F89E78
