

-----  
Midmark M11 - v1.0.5  
Total Cycles: 550

Name: RITTER2  
Sterilizer ID: V2313228  
11/15/2021 11:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.1 F 0.0  
2:00 189.1 F 3.9  
4:00 200.8 F 9.6  
6:00 220.5 F 6.5  
8:00 231.4 F 11.6  
10:00 235.1 F 9.9  
12:00 236.6 F 8.9  
14:00 250.7 F 16.5  
16:00 263.3 F 23.9  
18:00 271.3 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.4  
1:00 272.2 F 29.9  
2:00 271.3 F 29.5  
3:00 272.2 F 29.8  
4:00 271.2 F 29.4  
  
Min 271.2 F 29.2  
Max 272.4 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.4 F 0.0  
5:00 208.4 F 0.0  
10:00 221.8 F 0.0  
15:00 223.8 F 0.0  
20:00 223.8 F 0.0  
25:00 223.8 F 0.0  
30:00 223.6 F 0.0  
Drying Complete

FILLING: 2:27  
HEATING: 18:01  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:57:40  
CYCLE COMPLETE

11/15/2021 12:32 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02