
Midmark M11 - v1.0.0
Total Cycles: 4828

Name: RITTER 1
Sterilizer ID: V2178198
08/16/2024 03:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.5 F 0.0
2:00 178.4 F 2.0
4:00 186.0 F 6.0
6:00 208.6 F 10.8
8:00 219.7 F 4.6
10:00 228.7 F 9.6
12:00 235.2 F 9.3
14:00 240.9 F 11.6
16:00 238.6 F 9.8
18:00 251.8 F 16.2
20:00 262.8 F 22.5
22:00 272.7 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.4 F 29.0
2:00 271.9 F 28.5
3:00 271.2 F 28.2
4:00 272.3 F 29.0

Min 271.2 F 28.1
Max 272.4 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.1 F 0.0
5:00 214.8 F 0.0
10:00 215.3 F 0.0
15:00 221.5 F 0.0
20:00 230.1 F 0.0
25:00 228.1 F 0.0
30:00 224.3 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 22:25
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:01:36
CYCLE COMPLETE

08/16/2024 04:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____