

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000008  
7:41 08/01/2020

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:22  
PRESSURIZING 2:26  
STERILIZING 3:31  
136.5 C 223kPa 3:31  
136.2 C 222kPa 4:01  
136.4 C 223kPa 4:31  
136.4 C 223kPa 5:01  
136.1 C 221kPa 5:31  
136.4 C 224kPa 6:01  
136.4 C 222kPa 6:31  
136.3 C 222kPa 7:01  
136.3 C 222kPa 7:01  
VENTING 7:01  
AIR DRYING 7:49  
CYCLE COMPLETE 8:04

-----  
Digital Signature #  
5093471D3CAA066B  
-----