

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000125  
7:14 08/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:06  
PRESSURIZING 4:12  
STERILIZING 5:49  
136.1 C 221kPa 5:49  
136.3 C 224kPa 6:19  
136.4 C 222kPa 6:49  
136.3 C 224kPa 7:19  
136.3 C 223kPa 7:49  
136.4 C 224kPa 8:19  
136.3 C 224kPa 8:49  
136.4 C 225kPa 9:19  
136.4 C 225kPa 9:19  
VENTING 9:19  
AIR DRYING 10:19  
CYCLE COMPLETE 10:23  
-----

Digital Signature #  
90C875FC1DBAD63F  
-----