

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1544

Name: RITTER2  
Sterilizer ID: V2313228  
09/27/2022 12:41 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.0 F 0.0  
2:00 167.8 F 2.9  
4:00 182.5 F 7.3  
6:00 204.7 F 13.2  
8:00 224.7 F 8.7  
10:00 237.4 F 10.3  
12:00 239.8 F 12.4  
14:00 240.6 F 11.5  
16:00 254.4 F 18.6  
18:00 266.5 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.4  
1:00 272.5 F 30.0  
2:00 271.7 F 29.3  
3:00 272.6 F 30.2  
4:00 272.3 F 29.7  
  
Min 271.5 F 29.3  
Max 272.7 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.8 F 0.0  
5:00 198.2 F 0.0  
10:00 215.4 F 0.0  
15:00 217.6 F 0.0  
20:00 221.9 F 0.0  
25:00 224.7 F 0.0  
30:00 227.6 F 0.0  
Drying Complete

FILLING: 2:44  
HEATING: 19:29  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:59:26  
CYCLE COMPLETE

09/27/2022 01:40 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_