
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004640
13:36 02/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:32
STERILIZING 3:20
136.0 C 220kPa 3:20
136.2 C 223kPa 3:50
136.2 C 222kPa 4:20
136.4 C 225kPa 4:50
136.4 C 222kPa 5:20
136.4 C 226kPa 5:50
136.1 C 223kPa 6:20
136.2 C 224kPa 6:50
136.2 C 224kPa 6:50
VENTING 6:50
AIR DRYING 7:39
CYCLE COMPLETE 11:11

Digital Signature #
3D3933CFBF534AE4
