

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001539  
14:40 10/28/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:51  
CYCLE ABORTED 6:44  
WATER LEVEL LOW  
NOT STERILE  
END TIME 6:44  
-----

Digital Signature #  
379B39E8295DAC5B  
-----