
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000839
13:51 06/04/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:04
PRESSURIZING 6:53
STERILIZING 8:21
136.2 C 221kPa 8:21
136.4 C 223kPa 8:51
136.2 C 221kPa 9:21
136.3 C 222kPa 9:51
136.4 C 223kPa 10:21
136.3 C 220kPa 10:51
136.3 C 222kPa 11:21
136.4 C 224kPa 11:51
136.4 C 224kPa 11:51
VENTING 11:51
AIR DRYING 12:57
CYCLE COMPLETE 18:01

Digital Signature #
F8EC738C819DF7FB
