

-----  
Midmark M11 - v1.0.5  
Total Cycles: 95

Name: RITTER2  
Sterilizer ID: V2313228  
06/14/2021 11:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 119.1 F 0.0  
2:00 120.5 F 1.1  
4:00 135.3 F 3.7  
6:00 166.3 F 7.1  
8:00 186.5 F 11.5  
10:00 209.7 F 5.3  
12:00 221.3 F 9.8  
14:00 232.6 F 9.7  
16:00 237.8 F 11.9  
18:00 237.4 F 10.2  
20:00 249.7 F 16.3  
22:00 260.6 F 22.5  
24:00 270.6 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.8  
1:00 272.3 F 30.0  
2:00 272.1 F 30.5  
3:00 271.3 F 29.7  
4:00 272.1 F 30.3

Min 271.2 F 29.6  
Max 272.7 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.4 F 0.0  
5:00 215.9 F 0.0  
10:00 230.2 F 0.0  
15:00 226.5 F 0.0  
20:00 212.0 F 0.0  
25:00 190.3 F 0.0  
30:00 181.6 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 24:42  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 01:03:07  
CYCLE COMPLETE

06/14/2021 12:41 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01