
Midmark M11 - v1.0.0
Total Cycles: 1640

Name: RITTER 1
Sterilizer ID: V2178198
02/16/2022 10:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 104.5 F 0.0
2:00 108.5 F 0.0
4:00 133.4 F 0.5
6:00 160.0 F 2.4
8:00 188.2 F 5.7
10:00 202.2 F 10.0
12:00 219.4 F 6.7
14:00 227.4 F 8.3
16:00 238.5 F 13.0
18:00 238.3 F 10.0
20:00 236.7 F 8.5
22:00 248.8 F 14.5
24:00 260.7 F 21.1
26:00 271.4 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.3 F 28.5
2:00 272.0 F 28.6
3:00 271.9 F 28.3
4:00 272.4 F 28.9

Min 271.5 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.5 F 0.0
5:00 207.7 F 0.0
10:00 220.0 F 0.0
15:00 242.4 F 0.0
20:00 222.1 F 0.0
25:00 221.7 F 0.0
30:00 220.9 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 26:37
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:04:36
CYCLE COMPLETE

02/16/2022 11:20 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____