
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002570
7:53 13/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:50
PRESSURIZING 2:55
STERILIZING 4:06
136.4 C 224kPa 4:06
136.2 C 222kPa 4:36
136.4 C 224kPa 5:06
136.3 C 223kPa 5:36
136.2 C 222kPa 6:06
136.4 C 222kPa 6:36
136.4 C 224kPa 7:06
136.2 C 222kPa 7:36
136.2 C 222kPa 7:36
VENTING 7:36
AIR DRYING 8:25
CYCLE COMPLETE 17:29

Digital Signature #
84560AFE90B1F201
