

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1106

Name: RITTER 1  
Sterilizer ID: V2178198  
08/20/2021 01:07 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 133.4 F 0.0  
2:00 138.0 F 0.9  
4:00 165.8 F 2.9  
6:00 184.3 F 5.8  
8:00 199.2 F 9.8  
10:00 218.0 F 6.9  
12:00 224.4 F 8.2  
14:00 235.3 F 12.8  
16:00 235.2 F 9.6  
18:00 239.0 F 9.5  
20:00 246.3 F 13.6  
22:00 257.6 F 19.7  
24:00 268.2 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.7  
1:00 271.5 F 29.0  
2:00 272.6 F 29.3  
3:00 271.5 F 28.8  
4:00 271.2 F 28.9

Min 271.0 F 28.5  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.4 F 0.0  
5:00 204.6 F 0.0  
10:00 206.8 F 0.0  
15:00 198.3 F 0.0  
20:00 190.5 F 0.0  
25:00 208.0 F 0.0  
30:00 213.6 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 25:14  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 01:03:36  
CYCLE COMPLETE

08/20/2021 02:10 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_