

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2761

Name: RITTER 1  
Sterilizer ID: V2178198  
01/10/2023 04:20 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.5 F 0.0  
2:00 185.9 F 0.0  
4:00 197.4 F 4.8  
6:00 213.8 F 9.9  
8:00 225.6 F 6.8  
10:00 236.4 F 11.3  
12:00 235.9 F 9.0  
14:00 238.6 F 9.4  
16:00 249.6 F 15.0  
18:00 261.9 F 22.0  
20:00 272.7 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.5 F 29.1  
2:00 272.3 F 28.7  
3:00 271.5 F 28.2  
4:00 272.0 F 28.9

Min 271.4 F 28.1  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.8 F 0.0  
5:00 214.7 F 0.0  
10:00 230.9 F 0.0  
15:00 233.0 F 0.0  
20:00 232.5 F 0.0  
25:00 223.6 F 0.0  
30:00 218.4 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 20:25  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:58:37  
CYCLE COMPLETE

01/10/2023 05:18 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_