
Midmark M11 - v1.0.0
Total Cycles: 2836

Name: RITTER 1
Sterilizer ID: V2178198
01/31/2023 03:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.0 F 0.0
2:00 186.8 F 4.3
4:00 191.7 F 9.0
6:00 207.5 F 8.2
8:00 224.7 F 8.3
10:00 240.5 F 13.1
12:00 238.6 F 10.8
14:00 237.1 F 8.9
16:00 250.7 F 15.8
18:00 262.8 F 22.7
20:00 272.0 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 272.4 F 29.1
2:00 271.8 F 28.5
3:00 271.3 F 28.5
4:00 272.4 F 29.1

Min 271.1 F 28.2
Max 272.4 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.7 F 0.0
5:00 226.7 F 0.0
10:00 227.9 F 0.0
15:00 229.8 F 0.0
20:00 228.6 F 0.0
25:00 230.7 F 0.0
30:00 230.0 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 20:12
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:45
CYCLE COMPLETE

01/31/2023 04:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____