
Midmark M11 - v1.0.0
Total Cycles: 4203

Name: RITTER 1
Sterilizer ID: V2178198
01/15/2024 11:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.4 F 0.0
2:00 157.9 F 1.8
4:00 158.8 F 4.5
6:00 183.2 F 8.0
8:00 221.9 F 12.6
10:00 220.7 F 5.5
12:00 231.4 F 9.8
14:00 236.4 F 9.5
16:00 240.7 F 11.0
18:00 236.8 F 8.9
20:00 250.1 F 15.1
22:00 261.0 F 21.4
24:00 271.5 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.7 F 28.9
2:00 272.1 F 28.3
3:00 271.0 F 28.1
4:00 272.4 F 29.0

Min 271.0 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.5 F 0.0
5:00 228.6 F 0.0
10:00 222.5 F 0.0
15:00 228.6 F 0.0
20:00 228.1 F 0.0
25:00 229.0 F 0.0
30:00 229.3 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 24:42
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:03:26
CYCLE COMPLETE

01/15/2024 12:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____